



Original Article

Agreement on endoscopic ultrasonography-guided tissue specimens: Comparing a 20-G fine-needle biopsy to a 25-G fine-needle aspiration needle among academic and nonacademic pathologists

Priscilla A. van Riet, ¹ Djuna L. Cahen, ¹ Katharina Biermann, ¹ Bettina Hansen, ¹ Alberto Larghi, ³ Guido Rindi, ³ Giovanni Fellegara, ⁴ Paolo Arcidiacono, ⁵ Claudio Doglioni, ⁵ Nicola Liberta Decarli, ⁶ Julio Iglesias-Garcia, ⁷ Ihab Abdulkader, ⁷ Hector Lazare Iglesias, ⁷ Masayuki Kitano, ⁸ Takaaki Chikugo, ⁸ Satoru Yasukawa, ⁹ Hans van der Valk, ² Nam Quoc Nguyen, ¹⁰ Andrew Ruszkiewicz, ¹⁰ Marc Giovannini, ¹¹ Flora Poizat, ¹¹ Schalk van der Merwe, ¹² Tania Roskams, ¹² Erwin Santo, ¹³ Silvia Marmor, ¹³ Kenneth Chang, ¹⁴ Fritz Lin, ¹⁴ James Farrell, ¹⁵ Marie Robert, ¹⁵ Juan Carlos Bucobo, ¹⁶ Alan Heimann, ¹⁶ Francisco Baldaque-Silva, ¹⁷ Carlos Fernández Moro, ¹⁷ and Marco J. Bruno, ¹

¹Erasmus MC University Medical Center Rotterdam, ²Pathan Laboratory, Rotterdam, The Netherlands, ³Digestive Endoscopy Unit, University Hospital A. Gemelli, IRCCS, Rome, ⁴Department of Surgical Pathology, Diagnostic Center Italy, ⁵Vita Salute San Raffaele University, Milan, ⁶Santa Chiara Hospital, Trento, Italy, ⁷University Hospital of Santiago de Compostela, Santiago de Compostela, Spain, ⁸Kindai University, Osaka, ⁹Kyoto Prefectural University of Medicine, Kyoto, Japan, ¹⁰Royal Adelaide Hospital, Adelaide, Australia, ¹¹Institut Paoli-Calmettes, Marseilles, France, ¹²University Hospital Leuven, Leuven, Belgium, ¹³Tel Aviv Sourasky Medical Center, Tel Aviv, Israel, ¹⁴University of California, Irvine, ¹⁵Yale University School of Medicine, New Haven, ¹⁶Stony Brook University Hospital, New York, USA and ¹⁷Karolinska University Hospital, Huddinge, Sweden

Background and Aim: A recently carried out randomized controlled trial showed the benefit of a novel 20-G fineneedle biopsy (FNB) over a 25-G fine-needle aspiration (FNA) needle. The current study evaluated the reproducibility of these findings among expert academic and non-academic pathologists.

Methods: This study was a side-study of the ASPRO (ASpiration versus PROcore) study. Five centers retrieved 74 (59%) consecutive FNB and 51 (41%) FNA samples from the ASPRO study according to randomization; 64 (51%) pancreatic and 61 (49%) lymph node specimens. Samples were re-reviewed by five expert academic and five non-academic pathologists and rated in terms of sample quality and diagnosis. Ratings were compared between needles, expert academic and non-academic pathologists, target lesions, and cytology versus histological specimens.

Results: Besides a higher diagnostic accuracy, FNB also provided for a better agreement on diagnosing malignancy ($\kappa = 0.59 \text{ vs } \kappa = 0.76, P < 0.001$) and classification according to Bethesda ($\kappa = 0.45 \text{ vs } \kappa = 0.61, P < 0.001$). This equally applied for expert academic and non-academic pathologists and for pancreatic and lymph node specimens. Sample quality was also rated higher for FNB, but agreement ranged from poor ($\kappa = 0.04$) to fair ($\kappa = 0.55$). Histology provided better agreement than cytology, but only when a core specimen was obtained with FNB (P = 0.004 vs P = 0.432).

Conclusion: This study shows that the 20-G FNB outperforms the 25-G FNA needle in terms of diagnostic agreement, independent of the background and experience of the pathologist. This endorses use of the 20-G FNB needle in both expert and lower volume EUS centers.

Key words: FNA, FNB, interobserver agreement, pathology

Corresponding: Priscilla A. van Riet, Department of Gastroenterology and Hepatology, Erasmus MC University Medical Center Rotterdam, P.O Box 2040, 3015 CE Rotterdam, The Netherlands. Email: p.vanriet@live.nl Received 17 January 2019; accepted 17 April 2019.

© 2019 The Authors. Digestive Endoscopy published by John Wiley & Sons Australia, Ltd on behalf of Japan Gastroenterological Endoscopy Society.

This is an open access article under the terms of the Creative Commons Attribution-NonCommercial-NoDerivs License, which permits use and distribution in any medium, provided the original work is properly cited, the use is non-commercial and no modifications or adaptations are made.

INTRODUCTION

RADITIONALLY, ENDOSCOPIC ULTRASOUND (EUS)-guided tissue sampling has been carried out using a thin and flexible fine-needle aspiration (FNA) needle, which mainly yields individual cells (cytology) rather than histologically intact tissue fragments. Although diagnostic accuracy rates of FNA are fair, intact tissue fragments are preferred to enable identification of tumor invasion and allow for ancillary immunological and molecular testing; for example, in submucosal and neuroendocrine tumors. ¹⁻⁹ Furthermore, histology enables genetic profiling and a patient-tailored approach, which is becoming increasingly relevant in this era of personalized medicine. ¹⁰⁻¹⁴ The growing need for histology resulted in the introduction of fine-needle biopsy (FNB) needles.

So far, most studies reported an equal performance of FNA and FNB needles, ^{6-9,15} but, recently, two large randomized trials showed a significant diagnostic benefit of FNB. ^{16,17} One of these studies, the randomized controlled ASPRO (ASpiration *vs* PROcore) trial, was carried out in 13 EUS clinics, worldwide. ¹⁷ This study showed a diagnostic benefit of a novel 20-G FNB needle (ProCore; Cook Medical, Bloomington, IN, USA) over the widely used 25-G FNA needle (EchoTip Ultra; Cook Medical), irrespective of lesion type, size, and the number of passes carried out. However, general applicability of these findings cannot be warranted, as study participation was confined to expert centers only.

Ideally, the superiority of a diagnostic device is reproducible in expert and non-expert hands. Therefore, the present study compares the diagnostic agreement on samples obtained with the novel 20-G FNB to the 25-G FNA needle among expert academic pathologists and non-academic pathologists.

METHODS Study design

IN THE COURSE of the ASPRO trial (ClinicalTrials.gov: NCT02167074), 13 EUS centers randomized 608 consecutive patients with a solid pancreatic lesion, lymph node, or submucosal or other solid lesion to sampling with a 20-G FNB (ProCore; Cook Medical) or with a 25-G FNA needle (EchoTip Ultra; Cook Medical), between February 2015 and September 2016. Parameters regarding specimen characteristics and diagnostic accuracy were compared. Gold standard diagnosis was based on the prior ASPRO study¹⁷ either on pathological evaluation of the surgical resection specimens or on clinical follow up for at least 9 months when surgical resection was not indicated. Gold standard

diagnosis was recorded by the principal investigator of each of the participating centers.

For the present side-study, the first 125 pancreatic and lymph node cases that were enrolled in the ASPRO study were included. Samples from these cases were reassessed by five expert academic and five non-academic pathologists. Diagnosis of malignancy and quality scores were assessed, and agreement on these outcome measures was compared between the two needles and between academic and non-academic pathologists.

As our study was a clinical trial, all authors could access the study data and have reviewed and approved the final manuscript.

Center, pathologist and case selection

ASpiration versus ProCore study centers were invited to contribute to this study if they had collected at least 20 solid pancreatic and lymph node samples by April 2016, and their pathologist was trained to read both cytology and histology. Five ASPRO study centers fulfilled these criteria (Milan, Osaka-Savama, Rome, Rotterdam, and Santiago de Compostela). Each center was represented by the specialized 'academic' pathologist who was also involved in the original ASPRO study. This academic pathologist invited a 'non-academic' colleague from a local community practice hospital with a general clinical profile to participate. Expert academic pathologists had reviewed between 3000 and 40 000 EUS samples, including both FNA and FNB during their career, whereas the non-academics had a sample review track record of between 50 and 1000. Per case, the academic pathologists selected the minimum number of slides required to obtain a tissue diagnosis, including immunohistochemically stained slides, if available.

Endoscopic ultrasonography-guided tissue sampling

Endoscopic ultrasonography procedures were carried out with a convex array echoendoscope (either Pentax EG-3870 UTK or EG-3270UK; Pentax, Tokyo, Japan, or Olympus UTC 140/180/260; Olympus, Tokyo, Japan) as described in the ASPRO study. Three study sites had on-site pathological evaluation at their disposal (Milan, Rotterdam, and Santiago de Compostela).

Specimen processing

Tissue samples were preserved according to local practice. Cytological tissue samples were smeared onto glass slides and stained with (Diff Quick, RAL diagnostics) or hematoxylin and eosin (HE) staining (Milan, Osaka-Sayama, Rome). Remainder of the cytological specimens were collected in (CytoLyt, ThinPrep® CytoLyt Solution, Marlborough, MA, USA), saline (Osaka-Sayama), or formalin (Milan). Cell suspensions were processed into cell blocks using the (Cellient automated cell block system Hologic, Toronto, Canada) or agar technique (Milan, Rome, Santiago de Compostela). Osaka-Sayama did not further process cytology. Histology was collected in CytoLyt (Santiago de Compostela and Rotterdam) or formalin (Milan, Rome, Rotterdam, Osaka-Sayama). Samples collected in formalin were processed as paraffin blocks, sectioned at 3-4 microns, and stained with HE for morphological evaluation.

Review session

Cases were reviewed during a 2-day session at the Erasmus MC University Medical Center Rotterdam, the Netherlands in April 2016. Each expert academic pathologist presented the selected cases providing information on the patient's gender, age and relevant medical history, type of target lesion (lymph node or solid pancreatic lesion) and a summary of the EUS report. Pathologists were blinded for the final clinical and pathological outcomes. Slides were viewed simultaneously using a multi-headed light microscope, but assessed individually. Slides, representative of a case, were presented, including immunohistochemically stained slides, if available. Each pathologist reviewed all cases, including their own.

Outcome measures and definitions

Primary outcome measure was to compare the diagnostic agreement on samples obtained with the two needles. First, samples were assessed for malignancy (yes/no) and classified according to Bethesda (non-diagnostic, benign, atypimalignant). 18 cal/suspect malignancy, and pseudopapillary neoplasms were classified as malignant. Neuroendocrine and spindle cell tumors were classified as malignant only if they harbored high-grade dysplasia or an invasive component. Second, we evaluated whether diagnostic agreement for the two needles differed between expert academic and non-academic pathologists, between pancreatic and lymphatic lesions, and between specimens containing cytology and histology.

Furthermore, agreement on specimen quality parameters was assessed and compared between the two needles, and between expert academic and non-academic pathologists. The following quality parameters were scored: presence of artifacts, sample sufficiency, presence of target cells and tissue cores and suitability for additional analysis. Artifacts

were subdivided into five categories: poor fixation or drying artifacts, thick smears, blood clots, contamination with other cells (mesothelial, liver, gastric or intestinal epithelium), and other. Sample sufficiency was defined as the presence of sufficient target cells to obtain or exclude a certain diagnosis. Target cells were classified as less or more than 50%. Presence of tissue core was defined as the presence of a measurable microscopic cylinder containing target organ cells with preserved histological structure.

Last, we assessed if and to what extent pathologist's experience or specimen characteristics influenced diagnostic accuracy.

Statistical analysis

The sample size for this study was derived from Walter et al. 19 Given the availability of 10 observers (five academic and five non-academic pathologists), 50 samples are needed to be analyzed per needle type (50 \times 2 = 100 in total), given a one-sided alpha of 0.05, a power of 80%, a minimally acceptable interrater reliability of 0.6 for agreement on the presence of malignancy, and a minimal deviation from the interrater reliability of 0.2 between the two needles, n = 10. Interobserver agreement was calculated by the use of kappa statistics Fleiss' k-statistic and 95% confidence intervals (CI). Kappa-statistics were interpreted according to the convention of Landis and Koch: <0, no agreement; 0-0.20, slight agreement; 0.21-0.40, fair agreement; 0.41-0.60, moderate agreement; 0.61-0.80, substantial agreement; and 0.81-1.0, almost perfect agreement. The chi-squared test was used to compare the diagnostic agreement between the two study needles, academic and non-academic pathologists, target lesion types, and cytological and histological samples. Although all 10 observers assessed the samples for each of the outcome parameters, we reported only the average outcome per parameter. Last, univariate logistic regression analysis was applied to assess whether a pathologist's expertise and sample quality influenced diagnostic accuracy. Outcomes of this analysis are expressed as odds ratio (OR) with 95% CI. Statistical significance was defined as P < 0.05(two-tailed). Analyses were carried out using SAS version 9.4 (SAS Institute, Cary, NC, USA), and SPSS version 22, Statistical Package for the Social Sciences (SPSS Inc., Chicago, IL, USA).

RESULTS

Target lesion and procedure characteristics

TOTAL OF 125 samples were reviewed, of which 74 Awere collected by FNB (59%) and 51 by FNA (41%), with a mean of 2.8 needle passes. Sixty-four were solid

Table 1 Characteristics of samples obtained with the novel 20-G FNB vs the 25-G FNA needle among expert academic pathologists and non-academic pathologists

Variable, n (%)	All	FNB	FNA	
	(n = 125)	(n = 74)	(n = 51)	
Center of origin				
Rotterdam	33 (26)	23 (31)	10 (20)	
Rome	30 (24)	20 (27)	10 (20)	
Milan	22 (18)	11 (15)	11 (22)	
Santiago de	20 (16)	10 (14)	10 (20)	
Compostela				
Osaka-Sayama	20 (16)	10 (14)	10 (20)	
Target lesion				
Solid pancreatic	64 (51)	39 (53)	25 (49)	
lesion				
Lymph node	61 (49)	35 (47)	26 (51)	
Size (mm),	30.4 ± 1.3	31.5 ± 1.8	28.8 ± 1.8	
mean \pm SD				
Location of pancreation	c lesions			
Head	40 (62)	28 (72)	12 (48)	
Neck	5 (8)	2 (5)	3 (12)	
Corpus	12 (19)	7 (18)	5 (20)	
Tail	7 (11)	2 (5)	5 (20)	
Location of lymph no	des			
Mediastinal	21 (34)	14 (40)	7 (14)	
Abdominal	40 (66)	21 (60)	19 (73)	
Gold standard diagno				
Benign, normal	18 (14)	14 (19)	4 (8)	
tissue				
Sarcoidosis	1 (1)	O (O)	1 (2)	
Pancreatitis	2 (2)	1 (1)	1 (2)	
Leiomyoma	1 (1)	O (O)	1 (2)	
GIST, low grade	2 (2)	O (O)	2 (4)	
NET low grade	2 (2)	1 (1)	1 (2)	
NET high grade	4 (3)	3 (4)	1 (2)	
Leiomyosarcoma	1 (1)	1 (1)	0 (0)	
Solid	3 (2)	2 (3)	1 (2)	
pseudopapillary				
neoplasm				
Metastatic disease	13 (10)	6 (8)	7 (13)	
Malignant	11 (9)	5 (7)	6 (12)	
lymphoma				
Adenocarcinoma	67 (54)	41 (55)	26 (51)	

FNA, fine-needle aspiration; FNB, fine-needle biopsy; GIST, gastrointestinal stromal tumor; NET, neuroendocrine tumor.

pancreatic lesions (51%) and 61 were lymph nodes (49%) with a mean size of 30.4 ± 1.3 mm. Table 1 shows the case and sampling specifics. Techniques intended to increase the sample yield were applied in 94% of cases; suction with a syringe in 74 (63%), the slow-pull technique in 50 (37%), and a combination of the two in five (4%). The gold

standard diagnosis comprised 26 (21%) non-malignant cases and 99 (79%) malignant cases (Table 1). The gold-standard diagnosis was based on surgical resection specimens in 31 cases (25%).

Diagnostic accuracy and agreement

In line with the ASPRO study results, FNB samples provided higher accuracy than FNA for malignancy (88% vs 77%, P = 0.002) and classification according to Bethesda (76% vs 61%, P = 0.002). Regarding the primary question of diagnostic agreement, FNB samples provided better agreement on the presence of malignancy ($\kappa = 0.76 \ vs \ 0.59$, P < 0.001) and classification according to Bethesda ($\kappa = 0.61 \ vs \ 0.45$, P < 0.001, Table 2). This was true for both expert academic and non-academic pathologists (Table 2).

Assessment per target lesion showed that for lymph nodes, FNB provided higher agreement on the presence of malignancy and classification according to Bethesda. However, in pancreatic lesions, FNB only outperformed FNA for agreement on the Bethesda classification, not for the presence of malignancy (Table 3). When comparing histology to cytology, agreement on the presence of malignancy was better for histological samples, but agreement on the Bethesda classification was better for histological samples only if they had been obtained with the FNB needle (Table 4).

Specimen quality and agreement

Compared to FNA, FNB samples contained fewer artifacts (52% vs 45%, P = 0.007, Table 2), but agreement was low for both FNB ($\kappa = 0.10$; 95% CI 0.07-0.14) and FNA samples ($\kappa = 0.17$; 95% CI 0.13–0.21). Agreement did not differ between expert academic and non-academic pathologists for FNA (P = 0.132) or FNB (P = 0.212). Sample sufficiency for diagnosis, percentage of target cells, presence of tissue cores, and suitability for additional analysis were all better for FNB than for FNA, but, again, agreement on these parameters was poor ($\kappa = 0.04$) to fair ($\kappa = 0.55$, Table 2). As for the collection of histology, use of FNB provided histological samples more often than did FNA (70% vs 36%, P < 0.001, Table 2). Agreement on all of the above-mentioned quality parameters was highest for the expert academic pathologists. Furthermore, agreement among the expert academic pathologists was higher for FNB than for FNA specimens. In non-academic pathologists, however, FNB provided for better agreement than FNA only for the identification of tissue cores ($\kappa = 0.26 \text{ vs}$ 0.04, P < 0.001).

© 2019 The Authors. *Digestive Endoscopy* published by John Wiley & Sons Australia, Ltd on behalf of Japan Gastroenterological Endoscopy Society.

Table 2 Agreement on sample diagnosis and quality among the pathologist groups per needle type

Cases scored as	FNB $(n = 74)$	FNA $(n = 51)$	P-value
Malignant – no. (%)	47 (64)	27 (53)	<0.001
Agreement $-\kappa$ (95% CI)			
All	0.76 (0.73-0.79)	0.59 (0.55–0.63)	< 0.001
Expert academic	0.74 (0.66–0.81)	0.54 (0.45–0.62)	< 0.001
Non-academic	0.78 (0.71–0.85)	0.64 (0.55–0.72)	< 0.001
Bethesda classification – no. (%)			
Non-diagnostic	6 (8)	8 (16)	< 0.001
Benign	9 (12)	3 (6)	
Neoplastic	12 (16)	13 (26)	
Malignant	47 (64)	27 (53)	
Agreement $-\kappa$ (95% CI)			
All	0.61 (0.60–0.64)	0.45 (0.43–0.48)	< 0.001
Expert academic	0.62 (0.57–0.67)	0.43 (0.37–0.49)	< 0.001
Non-academic	0.59 (0.55-0.64)	0.46 (0.40–0.52)	< 0.001
Sufficient quality – no. (%)	67 (91)	40 (78)	< 0.001
Agreement $-\kappa$ (95% CI)			
All	0.49 (0.46-0.53)	0.48 (0.44–0.52)	0.366
Expert academic	0.50 (0.43-0.58)	0.33 (0.28–0.37)	< 0.001
Non-academic	0.42 (0.35-0.49)	0.46 (0.37–0.54)	0.358
Target cells ≥50% – no. (%)	50 (68)	29 (57)	< 0.001
Agreement $-\kappa$ (95% CI)			
All	0.31 (0.28-0.34)	0.38 (0.33–0.41)	< 0.001
Expert academic	0.33 (0.26–0.40)	0.55 (0.47–0.64)	< 0.001
Non-academic	0.27 (0.20-0.34)	0.33 (0.24–0.42)	0.127
Tissue core present – no. (%)	52 (70)	18 (35)	< 0.001
Agreement $-\kappa$ (95% CI)			
All	0.37 (0.34–0.41)	0.14 (0.10-0.18)	< 0.001
Expert academic	0.41 (0.34-0.48)	0.08 (0.00-0.16)	< 0.001
Non-academic	0.26 (0.19-0.33)	0.04 (-0.04 to 0.13)	< 0.001
Additional analysis possible – no. (%)	56 (76)	28 (55)	< 0.001
Agreement $-\kappa$ (95% CI)			
All	0.47 (0.43-0.50)	0.42 (0.38–0.46)	0.016
Expert academic	0.51 (0.44–0.58)	0.43 (0.34–0.51)	0.042
Non-academic	0.38 (0.30–0.45)	0.38 (0.29–0.47)	0.593

CI, confidence interval; FNA, fine-needle aspiration; FNB, fine-needle biopsy.

 Table 3
 Diagnostic agreement of FNA and FNB per target lesion

Scored variablesAgreement κ (95% CI)	FNB $(n = 74)$	FNA $(n = 51)$	P-value
Bethesda classification			
Pancreas	0.54 (0.51–0.58)	0.47 (0.43-0.52)	< 0.001
Lymph node	0.64 (0.61–0.67)	0.43 (0.39-0.47)	< 0.001
Presence of malignancy			
Pancreas	0.64 (0.59–0.69)	0.60 (0.54-0.66)	0.114
Lymph node	0.84 (0.79–0.89)	0.58 (0.52–0.63)	< 0.001

CI, confidence interval; FNA, fine-needle aspiration; FNB, fine-needle biopsy.

Factors affecting diagnostic accuracy

Besides the type of needle, other factors affecting EUSsample diagnosis are shown in Table 5. A pathologist's background (expert academic or non-academic) did not influence the diagnostic accuracy of either needle (P = 0.250). Presence of artifacts did have an effect, as this resulted in a lower diagnostic accuracy (P = 0.030). Last,

 Table 4
 Diagnostic agreement on cytological and histological specimens per needle type

Agreement κ (95% CI)	Cytology	Histology	P- value		
Bethesda classifi	Bethesda classification				
All samples	0.51 (0.49-0.52)	0.60 (0.59-0.61)	< 0.001		
(n = 121)					
FNA $(n = 47)$	0.49 (0.46-0.50)	0.52 (0.49-0.55)	0.432		
FNB $(n = 74)$	0.52 (0.49–0.54)	0.62 (0.61–0.63)	< 0.001		
Presence of malignancy					
All samples	0.76 (0.74–0.78)	0.97 (0.95–0.99)	< 0.001		
(n = 121)					
FNA $(n = 47)$	0.73 (0.71–0.76)	0.89 (0.86–0.92)	0.002		
FNB $(n = 74)$	0.78 (0.75–0.81)	0.99 (0.79–1.00)	<0.001		

 ${
m CI},$ confidence interval; ${
m FNA},$ fine-needle aspiration; ${
m FNB},$ fine-needle biopsy.

the presence of tissue cores significantly improved diagnostic accuracy (P = 0.003).

DISCUSSION

In ADDITION TO the previously reported diagnostic benefit of a novel 20-G FNB over a commonly used 25-G FNA needle, the present study shows that diagnostic agreement is also higher for FNB than for FNA samples. More importantly, agreement on FNB samples was higher among pathologists from different backgrounds (academic *vs* community practice) and with different levels of experience (high *vs* lower volume). The benefit of FNB equally applies to pancreatic and lymphatic target lesions. The finding that FNB samples were of better quality and harbored histology more often likely contributed to their superior diagnostic performance.

Most studies on EUS-needle devices have been carried out in expert high-volume centers. However, EUS-guided

tissue sampling is increasingly applied in lower-volume centers. So far, few studies have evaluated the reproducibility of EUS-FNA/FNB results. Moreover, most of these studies had a limited number of observers, concerned one type of target lesion, or were carried out in an academic practice only.^{20–24} Previous studies reported diagnostic agreement rates ranging from moderate to excellent for FNA ($\kappa = 0.45 - 0.89$) and FNB ($\kappa = 0.61 - 0.94$). Recently, a promising study aimed to validate a novel scoring system to further optimize diagnostic agreement among cytopathologists.²⁴ Unfortunately, despite the fact that observers were selected from tertiary centers, diagnostic agreement for pancreatic FNA specimens was still suboptimal ($\kappa = 0.56$). Compared to these agreement rates, the 20-G FNB needle performed well, especially when we take into account that pathologists from all over the world were included, academics and non-academics alike. The 20-G FNB needle may thus contribute to improve reproducibility of EUS-FNA/B diagnosis.

The first explanation for better agreement on FNB samples is its high tissue core rate, as the collection of histology rather than cytology was positively associated with higher agreement. This is supported by the finding that the cytological yield of FNB was also higher than for FNA, but only availability of tissue cores for histology, and not cytology, contributed to a better diagnostic accuracy. The importance of tissue core samples over cytological samples to reach a correct diagnosis when using an FNB needle has been previously described by others.²⁰ Compared to other FNB needles, the cytological yield of the current 20-G FNB needle was also high. 9,20,22,25-31 Whereas previous studies reported sufficient cellularity in 19%-52%, in the current study, this was 68%. The only device that provides higher histology and cytology rates is the 19-G needle, ^{22,25} which obtains cores in 88% of samples and an adequate amount of loose target cells in 91%. However, the reported clinical applicability of being

Table 5 Factors affecting diagnostic accuracy by univariate analysis

Diagnostic accuracyBethesda classification	UnivariateOR (95%CI)	P-value	Diagnostic accuracy for malignancy	UnivariateOR (95%CI)	P-value
Pathologist experience	Pathologist experience				
Expert academic	0.96 (0.82-1.12)	0.587	Academic	0.88 (0.70–1.10)	0.250
Non-academic			Non-academic		
Presence of artifacts			Presence of artifacts		
No	1.45 (1.22-1.74)	< 0.001	No	1.34 (1.03-1.75)	0.030
Yes			Yes		
Type of tissue			Type of tissue		
Histology	0.55 (0.32-0.94)	0.030	Histology	0.39 (0.21-0.72)	0.003
Cytology			Cytology		

CI, confidence interval; OR, odds ratio.

© 2019 The Authors. *Digestive Endoscopy* published by John Wiley & Sons Australia, Ltd on behalf of Japan Gastroenterological Endoscopy Society.

able to obtain tissue with the 19-G FNB needle (81%) is much lower than with the 20-G FNB needle (99%). Although the increased flexibility of the 20-G FNB needle is likely a major contributor to its better performance, other needle design adjustments may also have improved the tissue acquisition rate. 32,33

Another quality parameter that may have contributed to the high diagnostic agreement on samples obtained with FNB is a low artifact rate. Although artifacts do not necessarily decrease accuracy when abundant tissue is collected, previous studies have shown that they may hamper advanced genetic testing, for example.³⁴ Interestingly, agreement on the presence of artifacts was low for both needles (although slightly better for FNB than for FNA). This is in line with the fact that agreement on all sample quality parameters was rather low, similar to reports from others. 21,24 This may result from a lack of EUS-sample quality definitions. In the current study, we tried to minimize this limitation by using the predefined scoring system, as proposed by the Papanicolaou Society of Cytopathology in 2014.18

There are several limitations to our study. First, each academic pathologist brought and presented his or her own slides. Although they too were blinded for the final outcome, we cannot exclude recall bias. However, this only applied to a few cases per pathologist. Second, pathologists assessed samples individually, whereas, in daily practice, difficult cases are often discussed among colleagues. Therefore, interobserver agreements reported in the current study may underestimate real-life reproducibility. Third, our study involved pathologists from 10 centers from around the world, whereas previous studies were confined to no more than five centers from the same geographical region. In the absence of uniform guidelines for EUS-guided tissue sampling and processing, it is inevitable that there are geographical and institutional differences in the work-up of specimens. These differences may have resulted in slight differences in the appearance of specimens, which may have hampered interpretation by pathologists not familiar with certain preparation techniques. Last, it must be considered that all samples were collected by expert endosonographers. For an ideal assessment of the reproducibility of the outcome of the ASPRO study, the study should be repeated in low-volume centers, with less experienced endosonographers.

In conclusion, the present study demonstrates that the novel 20-G FNB needle outperforms the 25-G FNA needle in terms of diagnostic agreement, as its diagnostic superiority is not limited by the expertise and experience of the reviewing pathologist. Better sample quality and presence of histology seem to be the determinants responsible for the better diagnostic performance of the 20-G FNB needle. Together with the favorable accuracy rates from the previous ASPRO study, current findings advocate the use of the novel 20-G FNB needle in high- as well as in lower-volume EUS centers.

ACKNOWLEDGMENTS

THE AUTHORS WOULD like to thank Fabia Attili, Harry Aslanian, Adebowale Adeniran, John G. Lee, Mariachiara Petrone, Erwan Bories, Erez Scapa, Jonathan M. Buscaglia, Maoxin Wu. We received an unrestricted grant from Cook Medical Devices, Ireland.

CONFLICTS OF INTEREST

RISCILLA VAN RIET is a consultant for Cook Medical. Marco Bruno is a consultant and lectures for Cook Medical and a consultant and lectures for Boston Scientific. Kenneth Chang is a consultant for Cook Medical and Olympus. Masayuki Kitano received speaking honoraria from Olympus.

REFERENCES

- 1 Diaz Del Arco C, Esteban Lopez-Jamar JM, Ortega Medina L, Diaz Perez JA, Fernandez Acenero MJ. Fine-needle aspiration biopsy of pancreatic neuroendocrine tumors: correlation between Ki-67 index in cytological samples and clinical behavior. Diagn. Cytopathol. 2017; 45: 29-35.
- 2 Jhala N, Jhala D. Definitions in tissue acquisition: core biopsy, cell block, and beyond. Gastrointest. Endosc. Clin. N. Am. 2014; 24: 19-27.
- 3 Layfield LJ, Ehya H, Filie AC et al. Utilization of ancillary studies in the cytologic diagnosis of biliary and pancreatic lesions: The Papanicolaou Society of Cytopathology Guidelines. Diagn. Cytopathol. 2014; 11(Suppl 1): 4.
- 4 Ribeiro A, Vazquez-Sequeiros E, Wiersema LM, Wang KK, Clain JE, Wiersema MJ. EUS-guided fine-needle aspiration combined with flow cytometry and immunocytochemistry in the diagnosis of lymphoma. Gastrointest. Endosc. 2001; 53:
- 5 Kim GH, Cho YK, Kim EY et al. Comparison of 22-gauge aspiration needle with 22-gauge biopsy needle in endoscopic ultrasonography-guided subepithelial tumor sampling. Scand. J. Gastroenterol. 2014; 49: 347-54.
- 6 Khan MA, Grimm IS, Ali B et al. A meta-analysis of endoscopic ultrasound-fine-needle aspiration compared to endoscopic ultrasound-fine-needle biopsy: diagnostic yield and the value of onsite cytopathological assessment. Endosc Int Open 2017; 5: E363-75.
- 7 Yang Y, Li L, Qu C, Liang S, Zeng B, Luo Z. Endoscopic ultrasound-guided fine needle core biopsy for the diagnosis of

- pancreatic malignant lesions: a systematic review and metaanalysis. Sci. Rep. 2016; 6: 22978.
- 8 Oh HC, Kang H, Lee JY, Choi GJ, Choi JS. Diagnostic accuracy of 22/25-gauge core needle in endoscopic ultrasoundguided sampling: systematic review and meta-analysis. Korean J. Intern. Med. 2016; 31: 1073-83.
- 9 Bang JY, Hawes R, Varadarajulu S. A meta-analysis comparing ProCore and standard fine-needle aspiration needles for endoscopic ultrasound-guided tissue acquisition. Endoscopy 2015; 48: 339-49.
- 10 Larson BK, Tuli R, Jamil LH, Lo SK, Deng N, Hendifar AE. Utility of endoscopic ultrasound-guided biopsy for nextgeneration sequencing of pancreatic exocrine malignancies. Pancreas 2018: 47: 990-5.
- 11 Zhu P, Sun S. Endoscopic ultrasound pin-points the precision medicine for pancreatic cancer. Endosc Ultrasound. 2016; 5: 1-
- 12 Jameson JL, Longo DL. Precision medicine-personalized, problematic, and promising. N. Engl. J. Med. 2015; 372: 2229-34
- 13 Bournet B, Gayral M, Torrisani J, Selves J, Cordelier P, Buscail L. Role of endoscopic ultrasound in the molecular diagnosis of pancreatic cancer. World J. Gastroenterol. 2014; 20: 10758-68.
- 14 Meyer JM, Ginsburg GS. The path to personalized medicine. Curr. Opin. Chem. Biol. 2002; 6: 434-8.
- 15 Nagula S, Pourmand K, Aslanian H et al. Comparison of endoscopic ultrasound-fine-needle aspiration and endoscopic ultrasound-fine-needle biopsy for solid lesions in a multicenter, randomized trial. Clin. Gastroenterol. Hepatol. 2018; 16: 1307-13.e1.
- 16 Cheng B, Zhang Y, Chen Q et al. Analysis of fine-needle biopsy versus fine-needle aspiration in diagnosis of pancreatic and abdominal masses: a prospective, multicenter, randomized controlled trial. Clin. Gastroenterol. Hepatol. 2018; 16: 1314-
- 17 van Riet PA, Larghi A, Attili F et al. A multicenter randomized trial comparing a 25-gauge EUS fine-needle aspiration device with a 20-gauge EUS fine-needle biopsy device. Gastrointest. Endosc. 2019; 89: 329-39.
- 18 Pitman MB, Centeno BA, Ali SZ et al. Standardized terminology and nomenclature for pancreatobiliary cytology: The Papanicolaou Society of Cytopathology Guidelines. CytoJournal 2014; 11(Suppl 1): 3.
- 19 Walter SD, Eliasziw M, Donner A. Sample size and optimal designs for reliability studies. Stat. Med. 1998; 17: 101-10.
- 20 Attili F, Petrone G, Abdulkader I et al. Accuracy and interobserver agreement of the Procore 25 gauge needle for endoscopic ultrasound-guided tissue core biopsy. Dig. Liver Dis. 2015; 47: 943-9.
- 21 Mounzer R, Yen R, Marshall C et al. Interobserver agreement among cytopathologists in the evaluation of pancreatic endoscopic ultrasound-guided fine needle aspiration cytology specimens. Endosc Int Open 2016; 4: E812-19.
- 22 Petrone MC, Poley JW, Bonzini M et al. Interobserver agreement among pathologists regarding core tissue specimens

- obtained with a new endoscopic ultrasound histology needle; a prospective multicentre study in 50 cases. Histopathology 2013; 62: 602-8.
- 23 Larghi A, Correale L, Ricci R et al. Interobserver agreement and accuracy of preoperative endoscopic ultrasound-guided biopsy for histological grading of pancreatic cancer. Endoscopy 2015; 47: 308-14.
- 24 Marshall C, Mounzer R, Hall M et al. Suboptimal agreement among cytopathologists in diagnosis of malignancy based on endoscopic ultrasound needle aspirates of solid pancreatic lesions: a validation study. Clin. Gastroenterol. Hepatol. 2018; 16: 1114-22.e2.
- 25 Iglesias-Garcia J, Dominguez-Munoz JE, Abdulkader I et al. Influence of on-site cytopathology evaluation on the diagnostic accuracy of endoscopic ultrasound-guided fine needle aspiration (EUS-FNA) of solid pancreatic masses. Am. J. Gastroenterol. 2011; 106: 1705-10.
- 26 Alatawi A, Beuvon F, Grabar S et al. Comparison of 22G reverse-beveled versus standard needle for endoscopic ultrasound-guided sampling of solid pancreatic lesions. United European Gastroenterol J. 2015; 3: 343-52.
- 27 Bang JY, Hebert-Magee S, Trevino J, Ramesh J, Varadarajulu S. Randomized trial comparing the 22-gauge aspiration and 22gauge biopsy needles for EUS-guided sampling of solid pancreatic mass lesions. Gastrointest. Endosc. 2012; 76: 321-7.
- 28 Fabbri C, Fuccio L, Fornelli A et al. The presence of rapid onsite evaluation did not increase the adequacy and diagnostic accuracy of endoscopic ultrasound-guided tissue acquisition of solid pancreatic lesions with core needle. Surg. Endosc. 2017; **31**: 225-30.
- 29 Iwashita T, Nakai Y, Samarasena JB et al. High single-pass diagnostic yield of a new 25-gauge core biopsy needle for EUS-guided FNA biopsy in solid pancreatic lesions. Gastrointest. Endosc. 2013; 77: 909-15.
- 30 Kamata K, Kitano M, Yasukawa S et al. Histologic diagnosis of pancreatic masses using 25-gauge endoscopic ultrasound needles with and without a core trap: a multicenter randomized trial. Endoscopy 2016; 48: 632-8.
- Ramesh J, Bang JY, Hebert-Magee S et al. Randomized trial comparing the flexible 19G and 25G needles for endoscopic ultrasound-guided fine needle aspiration of solid pancreatic mass lesions. Pancreas 2015; 44: 128-33.
- 32 Polkowski M, Jenssen C, Kaye P et al. Technical aspects of endoscopic ultrasound (EUS)-guided sampling in gastroenterology: European Society of Gastrointestinal Endoscopy (ESGE) Technical Guideline - March 2017. Endoscopy 2017; **49**: 989-1006.
- 33 Polkowski M, Larghi A, Weynand B et al. Learning, techniques, and complications of endoscopic ultrasound (EUS)guided sampling in gastroenterology: European Society of Gastrointestinal Endoscopy (ESGE) Technical Guideline. Endoscopy 2012; 44: 190-206.
- 34 Ooi M, Phan A, Nguyen NQ. Future role of endoscopic ultrasound in personalized management of pancreatic cancer. Endosc Ultrasound. 2017; 6: 300-7.