

Circulation

JOURNAL OF THE AMERICAN HEART ASSOCIATION



Reference Chart Derived From Post–Stent-Implantation Intravascular Ultrasound Predictors of 6-Month Expected Restenosis on Quantitative Coronary Angiography

P. J. de Feyter, P. Kay, C. Disco and P. W. Serruys

Circulation 1999;100:1777-1783

Circulation is published by the American Heart Association, 7272 Greenville Avenue, Dallas, TX
72514

Copyright © 1999 American Heart Association. All rights reserved. Print ISSN: 0009-7322. Online
ISSN: 1524-4539

The online version of this article, along with updated information and services, is
located on the World Wide Web at:

<http://circ.ahajournals.org/cgi/content/full/100/17/1777>

Subscriptions: Information about subscribing to *Circulation* is online at
<http://circ.ahajournals.org/subscriptions/>

Permissions: Permissions & Rights Desk, Lippincott Williams & Wilkins, 351 West Camden
Street, Baltimore, MD 21202-2436. Phone 410-5280-4050. Fax: 410-528-8550. Email:
journalpermissions@lww.com

Reprints: Information about reprints can be found online at
<http://www.lww.com/static/html/reprints.html>

Reference Chart Derived From Post–Stent-Implantation Intravascular Ultrasound Predictors of 6-Month Expected Restenosis on Quantitative Coronary Angiography

P.J. de Feyter, MD, PhD; P. Kay, MD; C. Disco, MSC; P.W. Serruys, MD, PhD

Background—Intravascular ultrasound (IVUS)–guided stent implantation and the availability of a reference chart to predict the expected in-stent restenosis rate based on operator-dependent IVUS parameters may interactively facilitate optimal stent placement. The use of IVUS guidance protects against undue risks of dissection or rupture.

Methods and Results—IVUS-determined post–stent-implantation predictors of 6-month in-stent restenosis on quantitative coronary angiography (QCA) were identified by logistic regression analysis. These predictors were used to construct a reference chart that predicts the expected 6-month QCA restenosis rate. IVUS and QCA data were obtained from 3 registries (MUSIC [Multicenter Ultrasound Stenting in Coronaries study], WEST-II [West European Stent Trial II], and ESSEX [European Scimed Stent EXperience]) and 2 randomized in-stent restenosis trials (ERASER [Evaluation of ReoPro And Stenting to Eliminate Restenosis] and TRAPIST [TRApidil vs placebo to Prevent In-STent intimal hyperplasia]). In-stent restenosis was defined as luminal diameter stenosis >50% by QCA. IVUS predictors were minimum and mean in-stent area, stent length, and in-stent diameter. Multiple models were constructed with multivariate logistic regression analysis. The model containing minimum in-stent area and stent length best fit the Hosmer-Lemeshow goodness-of-fit test. This model was used to construct a reference chart to calculate the expected 6-month restenosis rate.

Conclusions—The expected 6-month in-stent restenosis rate after stent implantation for short lesions in relatively large vessels can be predicted by use of in-stent minimal area (which is inversely related to restenosis) and stent length (which is directly related to restenosis), both of which can be read from a simple reference chart. (*Circulation*. 1999;100:1777-1783.)

Key Words: ultrasonics ■ angiography ■ restenosis ■ stents

Intracoronary stent implantation has unequivocally been shown to reduce the frequency of 6-month in-stent restenosis in focal lesions in relatively large coronary vessels.^{1–4} Intravascular ultrasound (IVUS) has successfully been used to guide stent implantation and often results in a wider stented lumen, which may, according to the “bigger is better” principle, further reduce the restenosis rate.^{5–21} In the present study, we sought to identify post–stent-implantation IVUS predictors of in-stent restenosis. These parameters are partly operator dependent and are available online; thus, they can be adjusted by the operator during actual stent implantation to achieve the best possible result.

We performed a meta-analysis to determine final post–stent-implantation IVUS predictors of 6-month restenosis (diameter stenosis \geq 50%) by quantitative coronary angiography (QCA) after stent implantation in focal lesions and large vessels. Patients were selected from 3 stent registries (MUSIC [Multicenter Ultrasound Stenting in Coronaries study],¹⁴ WEST-II [West European Stent Trial II],²² and

ESSEX [European Scimed Stent EXperience]) and from 2 randomized trials investigating the efficacy of a pharmaceutical agent on in-stent restenosis (ERASER [Evaluation of ReoPro And Stenting to Eliminate Restenosis] and TRAPIST [TRApidil vs placebo to Prevent In-STent intimal hyperplasia]). The latter 2 trials did not demonstrate a significant difference between drugs and placebo, and thus for the present study, all patients groups were combined. The registries and trials have been considered together because the IVUS and QCA analyses used were performed in the same core laboratory.

The analysis was performed by a multistep approach. First, IVUS predictors for 6-month in-stent (QCA) restenosis were identified. Second, various univariate and multivariate models were constructed with these parameters to predict the observed restenosis. Third, models were selected that had the best fit on the Hosmer-Lemeshow goodness-of-fit test. Finally, from the best-fit models, the model with the most easily obtainable and clinically relevant parameters was selected,

Received December 22, 1998; revision received July 7, 1999; accepted July 12, 1999.

From University Hospital Rotterdam–Dijkzigt, Rotterdam, Netherlands.

Correspondence to P.J. de Feyter, MD, University Hospital Rotterdam–Dijkzigt, Thoraxcenter Bd 377, PO Box 2040, 3000 CA Rotterdam, Netherlands.

© 1999 American Heart Association, Inc.

Circulation is available at <http://www.circulationaha.org>

TABLE 1. Features of In-Stent Restenosis Registries and Randomized Trials

Trial	Stent	Patients Recruited in Study, n	Patients With IVUS at Baseline, n	Study Patients With IVUS at Baseline and QCA at Follow-Up, n	Type of Study	IVUS-Guided Stenting	Restenosis by QCA, %
Balloon-expandable stents							
MUSIC	P-S (15 mm)	161	151	140	Registry	Yes	9.8
WEST-II	Multi-Link (15 mm)	165	151	137	Registry	Yes	11.7
ERASER	P-S (15 mm)	215	190	162	Randomized	Yes	19.8
Self-expanding stents							
ESSEX	Radius (14 mm)	103	92	89	Registry	Yes	21.3
TRAPIST	Wallstent (29 mm)	312	274	245	Randomized	Yes	25.7
Total		956	858	773			

P-S indicates Palmaz-Schatz.

and a reference chart was constructed that predicted the expected angiographic in-stent restenosis rate after stent implantation of short lesions in relative large vessels.

Methods

Patients were selected from 3 stent registries and 2 randomized restenosis trials (Table 1). Only patients who underwent single-lesion treatment were enrolled. Various stent types were implanted: 2 balloon-expandable stents (15-mm Palmaz-Schatz, Cordis Corp; and 15-mm Multi-Link, Guidant/Advanced Cardiovascular Systems) and 2 self-expanding stents (14-mm Radius, Scimed Life Systems; and the Wallstent, Schneider Europe [various stent lengths]) were used. All registries and trials used intracoronary ultrasound as guidance for optimal stent implantation. Patients were included in the present study if they had both an adequate final post-stent-implantation IVUS examination and a 6-month coronary angiogram to establish the restenosis rate. Six-month in-stent restenosis was determined with QCA and was defined as an in-stent luminal diameter stenosis $\geq 50\%$. The restenosis rate in the above-mentioned studies varied from 9.8% to 25.7%.

Baseline clinical characteristics are shown in Table 2. Overall, there were only small differences between the various patient groups except for the severity of angina, which was less severe in the WEST-II and ESSEX registries, and the higher frequency of prior myocardial infarction in the ERASER and TRAPIST randomized trials.

Intravascular Ultrasound

Three different intracoronary ultrasound imaging devices were used. The CVIS (Microview) has a 30-MHz single-element bevelled transducer mounted on the end of a flexible shaft that rotates at 1800 rpm within either a 2.9F echotransparent long-monorail/common-distal-lumen imaging sheath or within a 3.2F short-monorail imaging sheath. The Hewlett-Packard/Boston Scientific Corp device incorpo-

rates a single-element 30-MHz bevelled monorail transducer that rotates at 1800 rpm within a 3.5F short-monorail imaging catheter. The Endosonics Europe device consists of 64 elements mounted at the tip, and the coaxial catheter allows either a monorail or an over-the-wire technique to be used.

IVUS imaging was performed during a motorized uniform pull-back speed of 0.5 mm/s in all patients. The video signal was recorded on a high-resolution super-VHS tape. The videotape was used to obtain manual online in-stent IVUS measurements to guide stent implantation. In-stent diameter (mm), in-stent minimum area (mm²), and lumen measurements at the 3- to 5-mm reference area proximal and distal to the stented segment were obtained. Stent apposition to the vessel wall and symmetrical stent expansion were reviewed over the entire stented segment. Optimal IVUS-guided stent implantation was attempted with the MUSIC criteria (complete stent apposition, symmetrical expansion, and adequate in-stent cross-sectional area), but it was left to the discretion of the operator to stop further attempts to achieve these criteria if additional dilation might pose a significant risk of dissection or rupture.¹⁴

Quantitative IVUS Measurements

A Microsoft Windows-based contour-detection program was used for automated 3D analysis of IVUS images. The contour-detection program was performed in 2 steps. First, 2 longitudinal sections were constructed from the 3D data set, and an automated contour-detection algorithm determined lumen-intima and media-adventitia boundaries of nonstented segments and lumen-stent boundary in stented segments. The contour-detection program uses a minimum-cost-based algorithm. These longitudinal contours were used to define regions of interest in the individual cross-sectional IVUS images. Next, the regions of interest were used to guide the second-final automated contour detection in these cross sections. A cursor on the longitudinal sections indicated the individual cross sections and allowed scrolling through the entire cross-sectional data set, to allow the technician to assess the automatically detected

TABLE 2. Baseline Clinical Characteristics

	MUSIC (n=151)	WEST-II (n=151)	ERASER (n=190)	ESSEX (n=92)	TRAPIST (n=274)
Age, y	60±10	60±10	60±9	62±9	60±9
Male sex, %	82	78	79	84	81
CCS III/IV, %	43	23	55	18	51
Diabetes mellitus, %	11	15	14	12	14
Prior MI, %	25	22	43	24	46
Prior CABG, %	2	4	7	3	5
Prior PTCA, %	9	12	13	25	17

CCS indicates Canadian classification angina; MI, myocardial infarction.

TABLE 3. Final IVUS Measurements After Stent Implantation

	MUSIC (n=151)	WEST-II (n=151)	ERASER (n=190)	ESSEX (n=92)	TRAPIST (n=274)
Minimum in-stent area, mm ²	7.85±2.0	7.39±1.9	7.04±2.0	6.96±2.0	6.17±1.9
Projected minimum lumen diameter, mm	2.77±0.4	2.78±0.4	2.67±0.4	2.67±0.4	2.59±0.4
Mean in-stent area, mm ²	9.45±2.4	8.95±2.2	8.71±2.3	8.53±2.3	7.83±2.1
Stent volume, mm ³	NA	154±75	168±113	128±58	227±131
Stent length, mm	NA*	17.1±6.6	19.5±11.9	14.8±4.2	28.6±13.2
Mean reference lumen area, mm ²	9.45±2.5	9.59±3.1	9.0±3.3	8.72±2.9	9.34±4.1

NA indicates not available in all patients.

*The length of the stent was 15.2±2.0, which was obtained from angiographic data.

contours. These user-interactive adjustments were made by “forcing” the contour through visually identified contour points, which then resulted in an upgrading of the entire data set (dynamic programming).

The automatic contour program allows for automated analysis of a maximum of 200 cross sections. The distance between 2 cross sections (slice thickness) therefore depends on the length of the segment of interest, with a minimum of 0.2 mm and a maximum of 0.5 mm.

In-stent IVUS measurements were as follows: in-stent minimum and mean diameter, in-stent mean and minimum area, stent volume, and stent length. Reference area measurements were the mean of the lumen area of the proximal and distal 5 mm of coronary lumen adjacent to the stent.

Volumes of total stent were calculated as

$$V = \sum_{i=1}^n A_i \cdot H$$

where V is volume, A is cross-sectional area of stent, H is thickness of the coronary slice, and n is the number of digitized cross sections encompassing the index volume. In-stent length was calculated as the number of images analyzed multiplied by the distance between 2 adjacent images.

The intraobserver and interobserver variabilities were <1.0%, and the SD of differences varied between 4% and 9.5%.^{16,17}

Quantitative Coronary Angiography

The minimal in-stent lumen diameter was determined on an end-diastolic frame by use of a computer-based Coronary Angiography Analysis System (CAAS II; Pie Medical). The edge-detection algorithm is based on the weighted sum of the first- and second-derivative functions applied to the digitized brightness silhouette.²³

The diameter function of the coronary artery lumen was determined by computing the shortest distance between the edge points of the right and left boundaries. The minimum lumen diameter was defined as the shortest distance between all measured left and right boundaries. The interpolated diameter was based on a computer estimation of the original lumen diameter, determined at the site of the minimum lumen diameter by taking into account the diameter function of the proximal and distal references. The diameter stenosis was derived from the measured minimum lumen diameter and the interpolated reference diameter.

Statistical Analysis

Statistical analysis was performed with SAS version 6.12. Quantitative data are given as mean±SD and qualitative data as frequencies.

IVUS parameters concerning volumetric measurements were not available in ≈67% of the patients selected from the MUSIC trial because a manual IVUS pullback was performed. Multiple IVUS parameters were tested by univariate logistic regression analysis to determine 6-month QCA restenosis predictors. With multivariate logistic regression analysis, multiple models containing IVUS parameters predictive of 6-month restenosis were constructed. These

models were then tested by use of the Hosmer-Lemeshow goodness-of-fit test to choose the most appropriate model. In the Hosmer-Lemeshow goodness-of-fit test, an estimated event probability is calculated from the observations by use of a model. These observations are then sorted in order of their estimated event probability and divided into ≈10 groups (g) of approximately equal size. The Hosmer-Lemeshow goodness-of-fit statistic is obtained by calculating the Pearson χ^2 test from the 2×g table of the observed and expected frequencies. The greater the P value, the better the model fits the data.

Results

The final post-stent-implantation IVUS measurements that were obtained in the various studies are tabulated in Table 3. In the MUSIC registry, the largest minimal in-stent cross-sectional area was achieved. The length of the implanted stent was longer, and consequently the in-stent volume was higher in the TRAPIST trial, which used Wallstents. The reference vessel diameter was smallest in the ESSEX registry. Univariate analysis of postprocedure final IVUS parameters showed that type of stent and all dimensional in-stent parameters except in-stent volume were significant predictors of restenosis (Table 4).

Multiple models were constructed with the IVUS parameters in univariate and multivariate analyses. When the Hosmer-Lemeshow goodness-of-fit test was used, 4 potentially appropriate models emerged (Table 5). One model was based on 1 parameter only (minimum in-stent area); the other 3 models were based on 2 parameters, and all 3 used stent length in combination with stent volume, mean stent area, or minimum stent area. Addition of any third parameter to a model did not significantly improve the predictive power. The combination of stent length and minimum in-stent area was chosen as the most easily obtainable and clinically

TABLE 4. IVUS Parameters After the Procedure: Univariate Analysis of Predictors of Restenosis Rate

Parameter	P
Minimum in-stent area	<0.001
Projected minimum lumen diameter	<0.001
Mean in-stent area	<0.001
Stent volume	0.174
Stent length	<0.001
Mean reference lumen area	0.020
Type of stent (Palmaz-Schatz, Radius, Multi-Link, Wallstent)	0.014

TABLE 5. Models (Univariate and Multivariate) Used to Predict Expected 6-Month QCA In-Stent Restenosis

Parameter	Coefficient	SE	P	Hosmer-Lemeshow Goodness-of-Fit Statistic (P)
Intercept	0.664
Minimum in-stent area	-0.325	0.06	<0.001	...
				7.60 (0.47)
Intercept	-2.136
Stent volume	-0.011	0.002	<0.001	...
Stent length	0.119	0.019	<0.001	...
				7.60 (0.47)
Intercept	0.009
Mean in-stent area	-0.264	0.054	<0.001	...
Stent length	0.031	0.008	<0.001	...
				6.41 (0.60)
Intercept	-0.199
Minimum in-stent area	-0.285	0.060	<0.001	...
Stent length	0.028	0.007	<0.001	...
				8.11 (0.42)

relevant parameters. The observed and expected restenosis rates of this model are depicted in Figure 1. The model remained accurate after the same analysis was performed with exclusion of Wallstent patients (who are known to have a higher restenosis rate), but the CIs were wider.

Next, a reference chart was developed (Table 6 and Figure 2) that can be used as a ready reference to estimate the expected in-stent restenosis rate with stent length and post-stent-implantation minimal in-stent area. The ranges of the 2 variables were divided into 10 groups each. The expected restenosis rate for the median of each range was calculated along with the 95% CIs for that particular value. The expected restenosis rate is given, as well as the CI.

The majority of observations were clustered around 15-mm-length stents and a minimum in-stent area of 8 mm² (Figure 3). Values in Table 6 that are marked by asterisks

were calculated by extrapolation from the model because there were no actual observations in that range. The reference chart in these ranges should be used with caution.

Discussion

In-stent restenosis remains a significant clinical problem because it is difficult to treat. IVUS-guided stent implantation has optimized stent expansion, and the initial higher lumen area achieved has been associated with a lower 6-month restenosis rate.^{14,18} We reasoned that construction of a reference chart based on simple, easily obtainable, clinically relevant IVUS post-stent-implantation parameters that predicts the expected 6-month restenosis rate would be clinically useful. First, online knowledge of an expected restenosis rate based on IVUS-determined post-stent-implantation lumen dimensions and stent length obtained while the procedure is being performed may serve as an extra stimulus to attempt to achieve the best possible result of stent implantation, whereas IVUS guidance may reduce the undue risks of dissection or vessel rupture.¹² Second, a priori knowledge of the expected restenosis rate based on the lesion length and vessel size, which are decisive in stent size and length selection, may be useful in counseling patients whether to select a percutaneous intervention or bypass surgery. Third, it may be helpful to discuss the probability of restenosis after stent implantation and accordingly plan closer monitoring of patients. Fourth, the reference chart may be helpful to plan and power future in-stent restenosis studies.

IVUS variables, including in-stent area, extent of preexisting plaque, smaller vessel size, stenting of total occlusions, and history of diabetes mellitus, as well as implantation of a long stent, have been shown to predict in-stent restenosis.^{10,20,21,24-26} In the present study, we found that 2 IVUS variables (minimum in-stent area and stent length) were strong predictors of 6-month restenosis. Using these 2 variables, it was possible for us to construct a reference chart that predicts the 6-month expected restenosis rate. Basically, the chart demonstrates that the widest in-stent area has the lowest restenosis rate, and the longer the implanted stent, the higher the restenosis rate. The achievement of maximal in-stent

Model - Minimum in-stent area and stent length




Figure 1. Reliability of model comparing observed vs expected 6-month restenosis rate (y axis) in various Hosmer-Lemeshow groups (x axis). Ideally, the observed and expected restenosis rates should be identical.

TABLE 6. Reference Chart

Stent Length, mm	Minimum In-Stent Area, mm ²									
	3.0–3.9	3.9–4.8	4.8–5.7	5.7–6.6	6.6–7.5	7.5–8.4	8.4–9.3	9.3–10.2	10.2–11.1	11.1–12.0
10–15	0.30 (0.22–0.40)	0.25 (0.19–0.32)	0.21 (0.16–0.26)	0.17 (0.13–0.21)	0.13 (0.11–0.17)	0.11 (0.08–0.14)	0.08 (0.06–0.12)	0.07 (0.04–0.10)	0.05 (0.03–0.09)	0.04 (0.02–0.08)
15–20	0.33 (0.25–0.42)	0.28 (0.22–0.34)	0.23 (0.19–0.28)	0.19 (0.16–0.22)	0.15 (0.12–0.18)	0.12 (0.09–0.16)	0.10 (0.07–0.13)	0.08 (0.05–0.12)	0.06 (0.04–0.10)	0.05 (0.02–0.09)
20–25	0.36 (0.28–0.46)	0.31 (0.25–0.37)	0.25 (0.21–0.30)	0.21 (0.18–0.24)	0.17 (0.14–0.20)	0.14 (0.11–0.17)	0.11 (0.09–0.15)	0.09 (0.06–0.13)	0.07 (0.04–0.11)	0.05* (0.03–0.10)
25–30	0.40 (0.31–0.49)	0.34 (0.27–0.41)	0.28 (0.24–0.33)	0.23 (0.20–0.27)	0.19 (0.16–0.23)	0.15 (0.12–0.20)	0.12 (0.09–0.17)	0.10 (0.06–0.15)	0.08 (0.05–0.13)	0.06 (0.03–0.11)
30–35	0.43 (0.34–0.53)	0.37 (0.30–0.44)	0.31 (0.26–0.38)	0.26 (0.22–0.31)	0.21 (0.17–0.26)	0.17 (0.13–0.22)	0.14 (0.10–0.19)	0.11 (0.07–0.17)	0.09* (0.05–0.15)	0.07* (0.04–0.13)
35–40	0.46 (0.36–0.57)	0.40 (0.32–0.49)	0.34 (0.28–0.41)	0.29 (0.23–0.35)	0.24 (0.19–0.29)	0.19 (0.14–0.25)	0.16 (0.11–0.22)	0.13 (0.08–0.19)	0.10* (0.06–0.17)	0.08* (0.04–0.15)
40–45	0.50 (0.39–0.61)	0.43* (0.34–0.53)	0.37 (0.29–0.46)	0.31 (0.25–0.39)	0.26 (0.20–0.34)	0.22* (0.15–0.29)	0.18 (0.12–0.25)	0.14* (0.09–0.22)	0.11 (0.06–0.19)	0.09* (0.05–0.17)
45–50	0.53* (0.41–0.65)	0.48 (0.36–0.58)	0.40 (0.31–0.50)	0.34 (0.26–0.44)	0.29 (0.21–0.38)	0.24 (0.17–0.33)	0.20 (0.13–0.29)	0.16* (0.09–0.25)	0.13 (0.07–0.22)	0.10* (0.05–0.19)
50–55	0.57* (0.43–0.69)	0.50 (0.38–0.62)	0.44 (0.33–0.55)	0.38 (0.28–0.49)	0.32* (0.22–0.43)	0.27* (0.18–0.38)	0.22* (0.14–0.33)	0.18* (0.10–0.29)	0.14* (0.08–0.26)	0.11* (0.05–0.22)
55–60	0.60* (0.45–0.73)	0.54* (0.40–0.67)	0.47* (0.34–0.60)	0.41* (0.29–0.54)	0.35 (0.24–0.48)	0.29* (0.19–0.42)	0.24 (0.15–0.38)	0.20* (0.11–0.33)	0.16* (0.08–0.29)	0.13* (0.06–0.26)

From IVUS-determined post-stent-implantation minimum in-stent area and stent length, one can extrapolate the expected 6-month QCA restenosis rate. The range of the 2 variables is divided into 10 groups each. The expected restenosis rate for the median of each range was calculated along with the 95% CI for this particular value (indicated in parentheses).

*There were no actual observations in this range; figures presented are calculated by extrapolation from the model.

dimensions expressed in terms of diameter, area, or volume is in keeping with the “bigger is better” theory and thus is associated with a lesser likelihood of reaching the expected restenosis rate.^{8,10,13,20,27,28}

The inverse relation of stent length and 6-month restenosis may be explained by the purely statistical fact that a longer stent has a higher chance of restenosis. However, a preexisting longer lesion, known to be associated with a higher restenosis rate, may be responsible. The inverse relation between stent length and restenosis may be taken as an argument for choosing a conservative stent-length implantation approach or “spot” stenting, but the efficacy of this approach should be tested in an appropriate randomized trial.

In the individual studies (Table 1), the restenosis rate was higher in self-expanding stents than balloon-expandable

stents, which suggests that the stent design may be associated with differences in restenosis rates. However, in the multivariate analysis, once we used the parameters of stent length and an area or volume measurement, stent design did not retain significance. It has been suggested that Wallstent implantation confers a higher restenosis rate. Therefore, we performed an analysis excluding Wallstent patients. The model was still accurate, but because of the lesser number of observations, it obviously is less reliable, with wider CIs.

We selected QCA at 6 months as our restenosis end point because the angiographic criterion of >50% diameter stenosis is a generally accepted criterion and can also be used, in contrast to IVUS-obtained restenosis parameters, in case of severe in-stent restenosis or a totally occluded stent.²⁹




Figure 2. Three-dimensional representation of restenosis rate (y axis) in relation to minimum in-stent area (x axis) and stent length (z axis).




Figure 3. Frequency distribution of number of observations (y axis) according to minimum in-stent area (x axis) and stent length (z axis).

Study Limitations

The study was performed in patients with focal lesions in relative large coronary vessels. Additional studies are needed to confirm that our results are also applicable in smaller vessels or for longer lesions.

Although this study was an analysis of >750 observations, the majority of these observations were clustered around a stent length of 15 mm and an in-stent minimum area of 8.0 mm², and many fewer observations were obtained for other stent lengths and in-stent cross-sectional areas, so that refinements may be expected from calculations obtained from a much larger data bank that includes a wide variety of in-stent cross-sectional areas and stent lengths.

The inclusion of stent length as a predictor of restenosis may be confounded by other factors, such as overlap of stent, implantation of long stents to cover initial long lesion, or long dissection, all of which may influence restenosis. However, the inclusion of these observations would more reliably reflect real-world stenting.

Currently, there is a wide variety of stent types commercially available. Our data are only applicable to the Palmaz-Schatz, Multi-Link, Radius, and Wallstents. Other stent designs (for instance, Gianturco-Roubin³⁰) may behave differently, and additional studies are needed to confirm the applicability of our reference chart to other stent types.

Conclusions

We constructed a reference chart that, by use of the IVUS-determined length of the implanted stent and the minimum in-stent area, predicted the 6-month QCA restenosis rate. Thus far, this reference chart is only applicable for stent implantation in short lesions in relatively large vessels. The reliability of this reference chart must be confirmed in a prospective study.

Acknowledgments

We pay tribute to all investigators who were involved in the MUSIC, WEST-II, and ESSEX registries and ERASER and TRAPIST randomized trials, without whose efforts this study would not have been possible. We wish to thank the core laboratory technicians and database managers at Cardialysis Rotterdam. Finally, we thank Claudia Sprenger de Rover for expert secretarial assistance.

References

- Fischman DL, Leon MB, Baim DS, Schatz RA, Savage MP, Penn I, Detre K, Veltri L, Ricci D, Nobuyoshi M, Cleman M, Heuser R, Almond D, Teirstein PS, Fish D, Colombo A, Brinker J, Moses J, Shaknovich A, Hirschfeld J, Bailey S, Ellis S, Rake R, Goldberg S. A randomized comparison of coronary-stent placement and balloon angioplasty in the treatment of coronary artery disease: Stent Restenosis Study Investigators. *N Engl J Med*. 1994;331:496–501.
- Serruys PW, de Jaegere P, Kiemeneij F, Macaya C, Rutsch W, Heyndrickx G, Emanuelsson H, Marco J, Legrand V, Materne P, Belardi J, Sigwart U, Colombo A, Goy JJ, van den Heuvel P, Delcan J, Morel MA. A comparison of balloon-expandable-stent implantation with balloon angioplasty in patients with coronary artery disease: Benestent Study Group. *N Engl J Med*. 1994; 331:489–495.
- Versaci F, Gaspardone A, Tomai F, Crea F, Chiariello L, Gioffre PA. A comparison of coronary-artery stenting with angioplasty for isolated stenosis of the proximal left anterior descending coronary artery. *N Engl J Med*. 1997;336:817–822.
- Serruys PW, van Hout B, Bonnier H, Legrand V, Garcia E, Macaya C, Sousa E, van der Giessen W, Colombo A, Seabra-Gomes R, Kiemeneij F, Ruygrok P, Ormiston J, Emanuelsson H, Fajadet J, Haude M, Klugmann S, Morel MA. Randomised comparison of implantation of heparin-coated stents with balloon angioplasty in selected patients with coronary artery disease (Benestent II). *Lancet*. 1998;352:673–681.
- Goldberg SL, Colombo A, Nakamura S, Almagor Y, Maiello L, Tobis JM. Benefit of intracoronary ultrasound in the deployment of Palmaz-Schatz stents. *J Am Coll Cardiol*. 1994;24:996–1003.
- Mudra H, Klauss V, Blasini R, Kroetz M, Rieber J, Regar E, Theisen K. Ultrasound guidance of Palmaz-Schatz intracoronary stenting with a combined intravascular ultrasound balloon catheter. *Circulation*. 1994; 90:1252–1261.
- Colombo A, Hall P, Nakamura S, Almagor Y, Maiello L, Martini G, Gaglione A, Goldberg SL, Tobis JM. Intracoronary stenting without anticoagulation accomplished with intravascular ultrasound guidance. *Circulation*. 1995;91:1676–1688.
- Dussailant GR, Mintz GS, Pichard AD, Kent KM, Satler LF, Popma JJ, Wong SC, Leon MB. Small stent size and intimal hyperplasia contribute to restenosis: a volumetric intravascular ultrasound analysis. *J Am Coll Cardiol*. 1995;26:720–724.
- Prati F, Di Mario C, Gil R, von Birgelen C, Camenzind E, Montauban van Swijndregt WJ, de Feyter PJ, Serruys PW, Roelandt JR. Usefulness of on-line three-dimensional reconstruction of intracoronary ultrasound for guidance of stent deployment. *Am J Cardiol*. 1996;77:455–461.
- Kastrati A, Schomig A, Elezi S, Schühlen H, Dirschinger J, Hadamitzky M, Wehinger A, Hausleiter J, Walter H, Neumann FJ. Predictive factors of restenosis after coronary stent placement. *J Am Coll Cardiol*. 1997; 30:1428–1436.
- von Birgelen C, Mintz GS, Nicosia A, Foley DP, van der Giessen WJ, Bruining N, Airrian SG, Roelandt JR, de Feyter PJ, Serruys PW. Electrocardiogram-gated intravascular ultrasound image acquisition after coronary stent deployment facilitates on-line three-dimensional reconstruction and automated lumen quantification. *J Am Coll Cardiol*. 1997; 30:436–443.
- Stone GW, Hodgson JM, St Goar FG, Frey A, Mudra H, Sheehan H, Linnemeier TJ. Improved procedural results of coronary angioplasty with intravascular ultrasound-guided balloon sizing: the CLOUT Pilot Trial: Clinical Outcomes With Ultrasound Trial (CLOUT) Investigators. *Circulation*. 1997;95:2044–2052.
- Savage MP, Fischman DL, Rake R, Leon MB, Schatz RA, Penn I, Nobuyoshi M, Moses J, Hirshfeld J, Heuser R, Baim D, Cleman M, Brinker J, Gebhardt S, Goldberg S. Efficacy of coronary stenting versus balloon angioplasty in small coronary arteries: Stent Restenosis Study (STRESS) Investigators. *J Am Coll Cardiol*. 1998;31:307–311.
- de Jaegere P, Mudra H, Figulla H, Almagor Y, Doucet S, Penn I, Colombo A, Hamm C, Bartorelli A, Rothman M, Nobuyoshi M, Yamaguchi T, Voudris V, DiMario C, Makovski S, Hausmann D, Rowe S, Rabinovich S, Sunamura M, van Es GA. Intravascular ultrasound-guided optimized stent deployment: immediate and 6 months clinical and angiographic results from the Multicenter Ultrasound Stenting in Coronaries Study (MUSIC study). *Eur Heart J*. 1998;19:1214–1223.
- Gorge G, Haude M, Ge J, Voegelé E, Gerber T, Rupprecht HJ, Meyer J, Erbel R. Intravascular ultrasound after low and high inflation pressure coronary artery stent implantation. *J Am Coll Cardiol*. 1995;26:725–730.
- von Birgelen C, Kutryk MJ, Gil R, Ozaki Y, Di Mario C, Roelandt JR, de Feyter PJ, Serruys PW. Quantification of the minimal luminal cross-sectional area after coronary stenting by two- and three-dimensional intravascular ultrasound versus edge detection and videodensitometry. *Am J Cardiol*. 1996;78:520–525.
- von Birgelen C, Gil R, Ruygrok P, Prati F, Di Mario C, van der Giessen WJ, de Feyter PJ, Serruys PW. Optimized expansion of the Wallstent compared with the Palmaz-Schatz stent: on-line observations with two- and three-dimensional intracoronary ultrasound after angiographic guidance. *Am Heart J*. 1996;131:1067–1075.
- Albiero R, Rau T, Schluter M, Di Mario C, Reimers B, Mathey DG, Tobis JM, Schofer J, Colombo A. Comparison of immediate and intermediate-term results of intravascular ultrasound versus angiography-guided Palmaz-Schatz stent implantation in matched lesions. *Circulation*. 1997;96:2997–3005.
- Kuntz RE, Gibson CM, Nobuyoshi M, Baim DS. Generalized model of restenosis after conventional balloon angioplasty, stenting and directional atherectomy. *J Am Coll Cardiol*. 1993;21:15–25.
- Hoffmann R, Mintz GS, Mehran R, Pichard AD, Kent KM, Satler LF, Popma JJ, Wu H, Leon MB. Intravascular ultrasound predictors of angiographic restenosis in lesions treated with Palmaz-Schatz stents. *J Am Coll Cardiol*. 1998;31:43–49.

21. Bauters C, Hubert E, Prat A, Bougrimi K, Van Belle E, McFadden EP, Amouyel P, Lablanche JM, Bertrand M. Predictors of restenosis after coronary stent implantation. *J Am Coll Cardiol*. 1998;31:1291–1298.
22. Serruys PW, van der Giessen WG, and the Investigators for the WEST-II Study. Clinical and angiographic results with the Multi-Link stent implanted under intravascular ultrasound guidance (West-2 Study). *J Invas Cardiol*. 1998;10(suppl B):20b–27b.
23. Haase J, Escaned J, van Swijndregt EM, Ozaki Y, Gronenschild E, Slager CJ, Serruys PW. Experimental validation of geometric and densitometric coronary measurements on the new generation Cardiovascular Angiography Analysis System (CAAS II). *Cathet Cardiovasc Diagn*. 1993;30:104–114.
24. Bermejo J, Botas J, Garcia E, Elizaga J, Osende J, Soriano J, Abeytua M, Delcan JL. Mechanisms of residual lumen stenosis after high-pressure stent implantation: a quantitative coronary angiography and intravascular ultrasound study. *Circulation*. 1998;98:112–118.
25. Hoffmann R, Mintz GS, Dussaillant GR, Popma JJ, Pichard AD, Satler LF, Kent KM, Griffin J, Leon MB. Patterns and mechanisms of in-stent restenosis: a serial intravascular ultrasound study. *Circulation*. 1996;94:1247–1254.
26. Kasaoka S, Tobis JM, Akiyama T, Reimers B, Di Mario C, Wong ND, Colombo A. Angiographic and intravascular ultrasound predictors of in-stent restenosis. *J Am Coll Cardiol*. 1998;32:1630–1635.
27. Elezi S, Kastrati A, Neumann FJ, Hadamitzky M, Dirschinger J, Schomig A. Vessel size and long-term outcome after coronary stent placement. *Circulation*. 1998;98:1875–1880.
28. Strauss BH, Serruys PW, de Scheerder IK, Tijssen JG, Bertrand ME, Puel J, Meier B, Kaufmann U, Stauffer JC, Rickards AF, Sigwart U. Relative risk analysis of angiographic predictors of restenosis within the coronary Wallstent. *Circulation*. 1991;84:1636–1643.
29. Foley DP, Melkert R, Serruys PW. Influence of coronary vessel size on renarrowing process and late angiographic outcome after successful balloon angioplasty. *Circulation*. 1994;90:1239–1251.
30. Topol EJ, Serruys PW. Frontiers in interventional cardiology. *Circulation*. 1998;98:1802–1820.